



innovation in
geotechnical
instrumentation

Calibration Record

RST Instruments Ltd., 11545 Kingston St., Maple Ridge, British Columbia, Canada V2X 0Z5
Tel: 604 540 1100 • Fax: 604 540 1005 • Toll Free: 1 800 665 5599 (North America only)
e-mail: info@rstinstruments.com • Website: www.rstinstruments.com

Vibrating Wire Piezometer

Customer: PT. Unggul Ejawantah Industri
Model: VW2100-0.35
Customer ID: PZ 15
Serial Number: VW57090
Mfg Number: P110321
Range: 350.0 kPa
Temperature: 23.0 °C
Barometric Pressure: 1028.8 millibars
Work Order Number: 220592
Cable Length: 100 meters
Cable Markings: 6567 m to 6667 m
Cable Colour Code: Red / Black (Coil) Green / White (Thermistor)
Cable Type: EL380004HDL
Thermistor Type: 3 kΩ

Applied Pressure (kPa)	First Reading (B units)	Second Reading (B units)	Average Reading (B units)	Calculated Linear (kPa)	Linearity Error (% FS)	Calculated Polynomial (kPa)	Polynomial Error (% FS)
0.0	9011	9011	9011	0.2	0.07	0.0	0.01
70.0	8364	8364	8364	69.9	-0.02	70.0	-0.01
140.0	7714	7715	7714	139.8	-0.05	140.0	-0.01
210.0	7063	7064	7064	209.9	-0.04	210.0	0.01
280.0	6412	6412	6412	280.0	0.00	280.1	0.02
350.0	5760	5761	5760	350.2	0.04	350.0	-0.01
Max. Error (%):					0.07		0.02

Linear Calibration Factor: CF = 1.0764E-01 kPa/B unit
Temperature Correction Factor: Tk = -4.8440E-03 kPa/°C rise

Polynomial Gage Factors:

A = -1.3660E-07 kPa/(B unit)² B = -1.0563E-01 kPa/B unit C = _____ kPa

Pressure is calculated with the following equations:

Linear: $P = CF(L_0 - L) - Tk(T_0 - T) + (S_0 - S)$

Polynomial: $P = A(L^2) + B(L) + C - Tk(T_0 - T) + (S_0 - S)$

Users must establish site zero readings for calculation purposes

Polynomial C = - [A(L₀²) + B(L₀)]

L₀, L = initial (installation) and current readings, in B units

T₀, T = initial (installation) and current temperature, in °C

S₀, S = initial (installation) and current barometric pressure readings, in kPa

B units = B scale output of VW 2102, VW 2104, VW 2106 and DT 2011 readouts

B units = Hz²/1000 ie: 1700 Hz = 2890 B units

	Date (dd/mm/yy)	VW Reading (B units)	Temperature (°C)	Baro (mbar)
Shipped Zero Readings:	29-Mar-19	9023	21.3	1027.5

This instrument has been calibrated using standards traceable to the NIST in compliance with ANSI Z540-1

Technician: M. Miftode

Date: 29-Mar-19



Document Number.: ELL0130K

